

التاريخ : 2025-10-21م

طرح عطاء رقم RFQ-H-11894

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين

الموضوع : **Stretcher**، حسب المواصفات .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكول الإسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق و تشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .
2. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.
3. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، و يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر دون تحمل اية مسؤولية تذكر في حالة تأخير التوريد ما يزيد عن اسبوعين .
4. تحديد فترة التوريد في عرض سعركم **وعدم ذلك يعنى أن التوريد فوري**، وايضاً أن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
5. **تسلم عروض الاسعار بالظرف المغلق فقط** ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني .
6. ضرورة الرد على العطاء سلباً أو ايجاباً.
7. الدفع حسب سياسات الدائرة المالية.
8. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار.
9. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع إعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، لا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل.
10. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
11. اخر موعد لتقديم عرض السعر هو يوم الثلاثاء الموافق 2025/10/28 الساعة العاشرة صباحاً (10:00AM).
12. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (0097092389687) داخلي (6400) أو من خلال البريد الإلكتروني **tender3@najah.edu فقط**.

مع فائق الاحترام ،،،،،

مدير دائرة اللوازم والمشتريات

إياد مكاي





التاريخ : 2025-10-21م

طرح عطاء رقم RFQ-H-11894

No	Item Code	Item	QTY	Unit	CODE NO. - وكيل حصري / موزع	Unit Price (Nis)	Total Price (Nis)
1	G-A- MDQ01150	Stretcher	5	PCS			
Total Price (NIS)							

- يرجى من حضرتكم تحديد مدة التوريد
- ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء، والالتزام بالعملة المحددة (شيكل)

- يرجى من حضرتكم الالتزام بتعبئة المعلومات في الجدول أدناه:

1	إسم الشركة
2	الختم
3	التوقيع والإسم
4	معلومات الإتصال
5	مدة التوريد
6	مدة الكفالة إن وجدت



Mobile Stretcher, Qty.: 5	Compliance Yes/No	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Should has ISO 13485 certification for medical device		
Manufacturer:		
Source and origin:		
Model:		
Mobile Stretcher, Specification:		
To be used in emergency, recovery and for patient transfer		
Adjustment for back section, trendelenburg, reverses trendelenburge, foot section and hight		
Should have hydraulic height adjustment from both sides of the trolley.		
Should have central braking system with steering facility.		
Should be equipped with heavy duty castors		
Should have bumpers at all the four corners and middle of the mattress base frame.		
Should have handle for transportation		
Should have place for fixing oxygen cylinder.		
Weight capacity not less than 200 Kg.		
Dimensions:		
Max length not less than 2000 mm		
Max width not less than 750 mm		
Height adjustable 550 – 900 mm approx.		
Trendelenburg up to 20 deg stepless.		
Reverse Trendelenburg up to 10 deg stepless.		
Complete with all accessories and consumables needed to work completely as specified		
The machine should be supplied with the following Accessories:		
1- Anti static hygienic mattress, 80mm with pull straps.		
2- Side Rails from both sides.		
3- IV rod qty. 1		
4- Foot end panel		



Optional to be priced separately:		
1- List price all accessories necessary to operate the solution as per the required specs listed above		
2- List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
3- Castor		
4- Gas spring		
Your offer should contain the following:		
1) Preventive maintenance schedule		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the product.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates		
8) The service and operational manuals must delivered as soft and hard copy with the units		
9) Guarantee for 10 years for spare parts supply		
Warranty: 3 years		

QW