



التاريخ : 26/08/2025 م

عطاء رقم – RFQ-H-11866

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء – المحترمين .

الموضوع : **Histology Stainer** ، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكال الاسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .

2. ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء مع ضرورة تعبئة جدول المواصفات الفنية.

3. ضرورة أن تكون الأجهزة المعروضة جديدة وليست مجدد .

4. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.

5. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، وفي حال تأخير التوريد مدة تزيد عن اسبوعين يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر مع تحميلكم فروق الأسعار .

6. الدفع خلال 90 يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .

7. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .

8. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي بالظرف المختوم فقط وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح أي منهما العرض الفني والعرض المالي.

9. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .

10. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار .

11. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .

12. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .

آخر موعد لتقديم عرض السعر هو يوم الاثنين الموافق 2025-09-08 م الساعة الواحدة ظهراً 01:00 pm .

13. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (97092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات

إياد المكاوي



التاريخ : 26/08/2025 م

طرح عطاء رقم – RFQ-H-11866


جدول الأسعار والكميات :

No	NNUH Code	Item	QTY	Unit	CODE #	وكيل / موزع	مدة التوريد	U.P (NIS)	T.P (NIS)
1	G-A-MDD01865	Histology Stainer حسب المواصفات المرفقة	1	Pcs					
-		Total Price (NIS)							

مع ضرورة ارفاق قائمة بأسعار المستهلكات الخاصة بالجهاز .

يرجى تعبئة الجدول التالي مع ضرورة الختم على جميع أوراق عرض السعر .

1	إسم الشركة
2	الختم
3	التوقيع والإسم
4	معلومات الإتصال
5	مدة التوريد
6	مدة الكفالة إن وجدت



جدول المواصفات الفنية :



Histology Stainer, Qty.: 1	Compliance Yes/No	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Should have CB test and EMC test (60601.1 and 60601.2)		
Should has ISO 13485 certification for medical devices		
IVD certificate		
Manufacturer:		
Source and origin:		
Model:		
Histology Stainer Specification:		
State of the art technology, the latest version, and the best technology from the manufacturer		
1. Applications		
Routine histology (Hematoxylin & Eosin, special stains)		
Cytology (Pap smear, Giemsa, others)		
Rapid H&E for frozen sections		
2. Capacity		
Reagent stations: 24 stations ($\leq$ 400 mL each)		
Washing stations: 2		
Cleaning station: 1		
Slide basket load: 30–40 slides per basket		
Multiple baskets supported for high-throughput operation (~200–400 slides/day)		
3. Programming		
20 independent staining programs		
Each program supports up to 20 steps		
Flexible immersion times: 1 second – 59 minutes 59 seconds		
User-friendly programming interface (touchscreen or control panel)		
4. Features & Benefits		
Heating/drying stations: Enables rapid processing of frozen sections and cytology slides		
Fume extraction system: Safe operation, reduces exposure to solvents		
Reagent rotation: Minimizes cross-contamination and reduces reagent waste		



Simple, robust design: Reliable, low-maintenance, suitable for routine lab workflows		
Continuous loading: Supports addition of slides during ongoing runs without interrupting the process		
5. Other Add-ons		
Remote monitoring / network connectivity		
Additional reagent trays for expanded protocols		
AC input: 220V, 50Hz		
Complete with all accessories and consumables needed to work completely as specified		
Training included:		
Operation training for the staff must be provided on-site		
Optional to be priced separately:		
List price all accessories necessary to operate the solution as per the required specs listed above		
List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
Your offer should contain the following:		
1) Preventive maintenance schedule		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the machine.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates and ISO 13485 certificates		
8) CB test and EMC test certificate		
9) The service and operational manuals must delivered as soft copy with the units		
10) Guarantee for 10 years for spare parts supply		
Warranty: 3 years, including the calibration, software upgrade, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.		
Preventive maintenance should be done as per the manufacturer's recommendations, all parts including PM kits to be replaced during the warranty period will be the responsibility of the supplier to provide on his own account. A schedule of the PM visits should be submitted to the hospital biomedical engineering dep.		
Hardware (service) keys, software passwords, website account or any means needed to grant the hospital engineer full access to all service functionalities are the right of the hospital and should be delivered with the equipment, and should as well be valid throughout the whole lifetime of machine.		



إعلان طرح عطاء رقم RFQ-H-11866

-- Histology Stainer --

- يعلن مستشفى النجاح الوطني الجامعي عن طرح عطاء (Histology Stainer)
- على السادة الشركات المتخصصة والراغبة بالمشاركة بالعطاء تقديم عروض أسعارها في موعد أقصاه يوم الإثنين الموافق 2025/09/08 الساعة الواحدة ظهراً 01:00 pm .
- ضرورة إرفاق نسخة من الرخصة التجارية والسيرة الذاتية.
- في حال وجود أي استفسار يرجى التواصل مع دائرة اللوازم والمشتريات هاتف رقم 2389685-02 تحويلة رقم 6400 أو من خلال البريد الإلكتروني Tender3@najah.edu

#	رقم العطاء	موضوع العطاء	ثمن نسخة العطاء	كفالة دخول عطاء	موعد الطرح	موعد الإغلاق	
						التاريخ	الساعة
1	RFQ-H-11866	<u>Histology Stainer</u>	مجانياً	-	2025/08/26	2025/09/08	AM 01:00