

التاريخ : 2024/07/15 م

طرح عطاء رقم – RFQ-H-11704

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين .

الموضوع : Portable Ultrasound ، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكال الاسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .
2. ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء .
3. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.
4. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، وفي حال تأخير التوريد مدة تزيد عن اسبوعين يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر مع تحميلكم فروق الأسعار .
5. الدفع خلال 90 يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .
6. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
7. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي بالظرف المختوم فقط وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح أي منهما العرض الفني والعرض المالي.
8. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .
9. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار .
10. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .
11. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
12. اخر موعد لتقديم عرض السعر هو يوم الإثنين الموافق 2024-07-22 الساعة الثالثة مساءً 15:00 pm .
13. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (+97092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات

إياد الكاوي



التاريخ : 15-07-2024 م

طرح عطاء رقم – RFQ-H-11704
جدول الأسعار والكميات :

No.	NNUH#	Item	QTY	Unit	Code + Brand name	Unit Price (Nis)	Total Price (Nis)
1	G-A-MDD01829	Portable Ultrasound حسب المواصفات المدرجة في الجدول أدناه صفحة رقم 5+4+3	1	Pcs			
-		Total Price (Nis)	-	-			



	إسم الشركة	1
	الختم	2
	التوقيع والإسم	3
	معلومات الإتصال	4
	مدة التوريد	5
	مدة الكفالة إن وجدت	6

جدول المواصفات الفنية :

Portable ultrasound, Qty.1	Compliance	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Should have CB test and EMC test (60601.1 and 60601.2)		
Manufacturer:		
Source and origin:		
Model:		
Portable Echocardiograph ultrasound Specification:		
Fully digital ultrasound system, compact portable ultrasound machine (Laptop Style)		
Heavy duty machine		
Color Doppler ultrasound for: Abdominal, obstetrical, small parts and general applications		
Automatic tissue optimization		
Imaging modes: B-Mode, M-Mode, Color M-Mode, Color Doppler, Power Doppler Imaging (PDI), PW Doppler, CW Doppler, Steerable Doppler, Tissue Harmonic Imaging (THI), Strain Imaging (SI), Strain Rate Imaging (SRI), 2D imaging, Color Flow Mode (CFM)		
High-definition (HD) digital black and white and colour liquid crystal display (LCD) monitor of at least 15 inches or more		
Display Resolution Not less than 1080 x 1280pixels		
Adjustable field of view		
Battery operated with battery life (run time) at least three hours		
Boot time less than 150 second		
The system must have archive capability for storage and retrieval of images and clips data		
Display Modes include: Duplex, Triplex, Split Screen, B/M & Dual B (B/B)		
Harmonic Imaging available on all probes		
System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations.		
Cine memory on all modes		
Integrated speakers, audio volume adjustment		
PAN/ZOOM: Realtime image, frozen / stored image		
Post Processing technique on the Frozen and stored images		
Multi - format Report		
Frame Rate: Greater than 700 fps		

Image recording capability		
Contrast adjustment (gain control)		
Beam Steer: shifts beam into needle for better needle visualization		
Image grey scale: 256 shades of grey		
The system shall process a dynamic range that is at least 150db.		
With two ports for probe		
System should support transducer technologies: Convex Array and Linear Array		
The system must display at a maximum depth of 30cm		
The system and probes should be strong enough to survive from any type of collision or accidental drop. Please send any certifications proving drop test		
Data Transfer facility should be available as standard to transfer images		
Archiving:		
With integrated HDD (500GB), integrated DVD-R/W & 2 USB (2.0) connectors		
Integrated image and patient data management package		
Export in BMP, JPEG, TIFF, DICOM, MOV and AVI format		
Ability to export report in PDF format		
Printing reports and images directly to printer		
Ability to be connected to HIS and RIS		
The system shall support the all DICOM functionality, Storage, Print, and Work List, also ready to connect to PACS		
Complete with all accessories and consumables needed to work completely as specified		
Training included:		
Operation training for the staff must be provided on-site		
Complete with the following accessories:		
Convex Probe 1 – 5 MHz		
Linear Probe 4 – 15 MHz		
Battery		
Optional to be priced separately:		
List price all accessories necessary to operate the solution as per the required specs listed above		

List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
Convex Probe 1 – 5 MHz		
Linear Probe 4 – 15 MHz		
Trolley		
Battery		

Your offer should contain the following:		
1) Preventive maintenance schedule		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the product.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates		
8) CB test and EMC test certificate		
9) The service and operational manuals must delivered as soft copy with the units		
10) Guarantee for 10 years for spare parts supply		

Warranty: 3 years, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations		
---	--	--



طرح عطاء

يعلن مستشفى النجاح الوطني الجامعي عن طرح عطاء رقم RFQ-H-11704 ، على الشركات المتخصصة والراغبة بالمشاركة بالعطاء تقديم عروض اسعرها في موعد اقصاه يوم الإثنين الموافق 22/07/2024 الساعة الثالثة 03:00 PM .

#	رقم العطاء	موضوع العطاء	نص نسخة العطاء	موعد الطرح	وقت الإغلاق	
					التاريخ	الساعة
1	RFQ-H-11704	Portable Ultrasound	مجانياً	2024/07/15م	PM 03:00	

حسب الشروط التالية:

1. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .

على الراغبين بالحصول على نسخة العطاء مراجعة دائرة اللوازم والمشتريات مكتب رقم 6400 / مبنى الإدارة أو عبر الإيميل Tender3@najah.edu وذلك خلال ساعات الدوام الرسمية .

للاستفسال يرجى الاتصال على:

لأية استفسارات أخرى يرجى مراجعة دائرة اللوازم والمشتريات هاتف رقم 970-09-2331471 / فاكس: +970-9-2389685 .

مستشفى النجاح الوطني الجامعي