

التاريخ : 24/12/2025 م

طرح عطاء رقم – RFQ-H-11922

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين .

الموضوع : أجهزة طبية ، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكل الاسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .

2. ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء .3. ضرورة أن تكون الأجهزة جديدة وليست مجددة .

4. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.

5. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، وفي حال تأخير التوريد مدة تزيد عن اسبوعين يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر مع تحميلكم فروق الأسعار .

6. الدفع خلال 90 يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .

7. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .

8. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي

بالظرف المختوم فقط وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني

أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح اي منهما العرض الفني والعرض المالي.

9. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .

10. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار .

11. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .

12. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .

13. آخر موعد لتقديم عرض السعر هو يوم الأحد الموافق 2025-01-11 م الساعة الواحدة ظهراً 01:00 PM .

14. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (+97092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات

إياد المكاوي

12600861+12600872+12600871

Page 1

طرح عطاء رقم – RFQ-H-11922

جدول الأسعار والكميات :

No .	NNUH Code	Item	QTY	Unit	CODE #	وكيل / موزع	مدة التوريد	U.P (NIS)	T.P (NIS)
1	G-A-MDD00008	Patient Warmer	2	Pcs					
2	G-A-MDD00017	Air Mattress	8	Pcs					
3	G-A-MDD00131	Compression Device	8	Pcs					
-	Total Price (NIS)								

يرجى تعبئة الجدول التالي مع ضرورة الختم على جميع أوراق عرض السعر .

1	إسم الشركة
2	الختم
3	التوقيع والإسم
4	معلومات الإتصال
5	مدة التوريد
6	مدة الكفالة إن وجدت


جدول المواصفات الفنية :




Item 1 Patient warming unit, Qty: 2	Compliance Yes/No	Deviation
Good brand name with FDA or CE		
Should has CB test and EMC test (60601.1 and 60601.2)		
Manufacturer:		
Source and origin:		
Model:		
Patient warming unit Specification:		
Mobile air patient warming unit		
Buit in heigh efficiency blower		
Adjustable temperature setting value		
Temperature accuracy +- 1		
Continuous operation		
Quiet, adjustable warming options		
Alarm for over temperature and automatic cut off of heating and air supply		
AC input: 220V, 50Hz		
Complete with all accessories and consumables needed to work completely as specified		
Training included:		
Operation training for the Hospital nursing team must be provided on-site		
Optional to be priced separately:		
1- List price all accessories necessary to operate the solution as per the required specs listed above		
2- List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
3- Filtter		
4- Adult fullbody blanket		
5- Pediatric fullbody blanket		

Your offer should contain the following:

1) Preventive maintenance schedule



2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the product.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates		
8) CB test and EMC test certificate		
9) The service and operational manuals must delivered as soft and hard copy with the units		
10) Guarantee for 10 years for spare parts supply		

Warranty: 1 year		
------------------	--	---

Air Mattress With Pump Qty.: 8	Compliance	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Should have CB test and EMC test (60601.1 and 60601.2)		
Should has ISO 13485 certification for medical devices		
Manufacturer:		
Source and origin:		
Model:		
Air Mattress With Pump Specification:		
State of the art technology, the latest version, and the best technology from the manufacturer		
Used for ICU patients for the prevention and treatment of pressure ulcers up to grade III according to EPUAP		
Cells technology for maximum comfort and effectiveness		
High quality cell material		
Quiet pump technology (low noise level)		
The system is suitable for patients up to 200 kg		
CPR emergency		
The compressors automatically switch off when it reached the pre set pressure		
Cycle modes: Alternating pressure cycle (APT) Static mode (CLP) MAX Transport mode		
Built in bracket for mounting easily to bed		
Pump type compressor		
Quick release connector		
The connector between pump and mattress should be connect from under the pump to prevent broken due to staff movement around the patient bed		
Length of mattress:180cm to 200cm. (Approximately)		
Width of mattress: 80cm to 90cm. (Approximately)		
Thickness of mattress: 15cm to 17.5cm. (Approximately)		
The mattress base should be of heavy duty and should be of non skid material		
The mattress cover must be able to withstand sterilizing materials		
Alarmfunctions (visual and audible): low pressure, power failure alarm		

Is the air pump need special tool to reset the pumps alarms?		
AC input: 220V, 50Hz		
Complete with all accessories and consumables needed to work completely as specified		
Training included:		
Operation training for the doctor and nursing team must be provided on-site		
Complete with the following accessories:		
Mattress		
Power cord		
Optional to be priced separately:		
List price all accessories necessary to operate the solution as per the required specs listed above		
List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
Mattress cover		
Your offer should contain the following:		
1) Preventive maintenance schedule		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the product.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates		
8) CB test and EMC test certificate		
9) The service and operational manuals must delivered as soft copy with the units		
10) Guarantee for 10 years for spare parts supply		
Warranty: 3 years, including the calibration, software upgrade, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.		
Preventive maintenance should be done as per the manufacturer's recommendations, all parts including PM kits to be replaced during the warranty period will be the responsibility of the supplier to provide on his own account. A schedule of the PM visits should be submitted to the hospital biomedical engineering dep.		
Hardware (service) keys, software passwords, website account or any means needed to grant the hospital engineer full access to all service functionalities are the right of the hospital and should be delivered with the equipment, and should as well be valid throughout the whole lifetime of machine.		

