

التاريخ : 2024/10/14 م

طرح عطاء رقم -- RFQ-H-11735


السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المترمين .

الموضوع: **Fully Automated Blood Bank Analyzer** مقابل شراء المواد التشغيلية الخاصة به على نظام ال **Lease** ، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الأسعار للمستهلكات اللازمة لتشغيل الجهاز بالشيكل الاسرائيلي وبشكل واضح وتشمل قيمة الضريبة المضافة.
2. ضرورة إرفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر.
3. أن يحتوي عرض السعر قائمة بجميع الفحوصات التي يمكن إجراؤها على الجهاز خصوصاً الفحوصات المرفقة في جدول العطاء بالإضافة إلى أسعار وأحجام جميع المستهلكات اللازمة لتشغيل الجهاز.
4. الإلتزام بتأمين مستهلكات الجهاز دون أي تأخير يؤدي إلى تعطيل العمل وتثبيت أسعارها لمدة 5 سنوات.
5. الإلتزام بتأمين صيانة الجهاز الدورية والطارئة وحسب تعليمات الشركة المصنعة مجاناً ولمدة 5 سنوات وتشمل قطع الغيار وأجور العمل.
6. تؤول ملكية جهاز ال **Fully Automated Blood Bank Analyzer** لمستشفى النجاح الوطني الجامعي بعد إنقضاء مدة خمسة سنوات على إستخدام الجهاز .
7. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .
8. الإلتزام بتدريب العاملين على الجهاز .
9. تلتزم الشركة الراغبة بالمشاركة بتخفيض سعر المستلزمات إذا تم تخفيض أسعارها عالمياً .
10. ضرورة أن يشمل عرض السعر عرضاً فنياً و عرضاً مالياً وأن يتم توضيح (**سعة التعبئة**) لكل بند يتم تسعيره .
11. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
12. تسلم عروض الأسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي **بالظرف المختوم فقط** وتوضع في صندوق العطاءات .
13. المستشفى غير ملزمة بأقل الأسعار .
14. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .
15. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
16. اخر موعد لتقديم عرض السعر هو يوم الثلاثاء الموافق 2024-10-22 الساعة الثالثة مساءً **15:00 pm** .
17. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (+97092389687) داخلي (6400) أو من خلال البريد الالكتروني **tender3@najah.edu** .

مدير دائرة اللوازم والمشتريات
إياد المكايوي



التاريخ : 14-10-2024 م

طرح عطاء رقم – RFQ-H-11735

جدول الأسعار والكميات :

No.	NNUH#	Item	QTY	Unit	Code + Brand name	Unit Price (Nis)	Total Price (Nis)
1	G-A-MDQ01336	Fully Automated Blood Bank Analyzer . حسب المواصفات الفنية المدرجة في الجدول صفحة 3+4+5	1	Pcs			
-		Total Price (Nis)	-	-			

حسب المواصفات الفنية المدرجة في الورقة رقم 3

	إسم الشركة	1
	الختم	2
	التوقيع والإسم	3
	معلومات الإتصال	4
	مدة التوريد	5
	مدة الكفالة إن وجدت	6

Automated Blood Bnak Analyzer, Qty.: 1	Compliance Yes/No	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Should has CB test and EMC test (60601.1 and 60601.2)		
Should has ISO 13485 certification for medical devices		
Manufacturer:		
Source and origin:		
Model:		
On Lease Contract with Pack-up System (same model or semi-automated system)		
Automated blood bnak analyzer specifications:		
State of the art technology, the latest version, and the best technology from the manufacturer		
Fully automated, continuous operation with minimal human intervention, automated sample loading, testing, and result interpretation		
Throughput capacity arround 120 tests / hour		
User-friendly touchscreen display		
STAT sample handling		
Compatible with whole blood, serum, and plasma samples		
The analyzer must have a design that minimizes carryover between samples		
On-board reagent stability: minimum 7 days		
Automated management of reagent levels with real-time monitoring and alerts for reagent replacement		
Integrated reagent barcode system for tracking and validation		
Storage of test results with the ability to generate detailed reports		
Full audit trail for results, quality control, and reagent use		
Bidirectional interface with hospital information systems (HIS)		



Immunoematology Testing Capabilities		
1. Blood Typing:		
a- ABO and Rh(D) typing with high specificity and sensitivity		
b- Capable of detecting weak D and partial D variants		
2. Antibody Screening:		
a- Automated screening for unexpected antibodies using indirect antiglobulin testing (IAT)		
b- Multiplex screening assays for higher sensitivity		
3. Crossmatching:		
a- Immediate Spin Crossmatch: Provides rapid compatibility assessment		
b- Antiglobulin Crossmatch: Comprehensive crossmatch for ensuring transfusion safety		
c- Automated result interpretation with clear indications of compatibility / incompatibility		
4. Antibody Identification:		
a- Ability to identify specific antibodies through panel testing		
b- Supports identification of clinically significant antibodies (e.g., anti-Kell, anti-Duffy)		
5. Coombs Testing:		
a- Direct Coombs Test: Detects antibodies bound to red cells in vivo		
b- Indirect Coombs Test: Detects free antibodies in the patient's serum		
6. Automated Antibody Titration:		
a- Fully Automated Titration Process: The analyzer must support automated antibody titration with minimal manual intervention		
7. Elution Testing:		
a- Automated elution of antibodies from sensitized red blood cells for detailed antibody investigation		
8. Cell Suspension Preparation:		
a- Automated preparation of standardized red cell suspensions for all testing		
b- Consistent cell concentration for reproducible results		
Advanced diagnostic tools for early detection of operational issues, flagging and troubleshooting alerts for handling errors, insufficient samples, or reagent issues		

JM

Automated or semi-automated calibration procedures with minimal technician intervention, with alert systems for when calibration is due		
Automated QC testing with the ability to schedule daily, weekly, and monthly checks, built-in quality control for reagents, including detection of lot variations		
Automatic waste management system		
Power supply: 230V, 50 Hz		
Specify the device dimension (HxWxD):		
Complete with all accessories and consumables needed to work completely as specified		
Training included:		
Operation training for the lab department team must be provided on-site		
Your offer should contain the following:		
1) Preventive maintenance schedule		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the product.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates		
8) CB test and EMC test certificate		
9) The service and operational manuals must delivered as soft and hard copy with the units		
10) Guarantee for 10 years for spare parts supply		

Preventive maintenance should be done as per the manufacturer's recommendations all parts including PM kits to be replaced during the lease period will be the responsibility of the supplier to provide on his own account. A schedule of the PM visits should be submitted to the hospital biomedical engineering dep.		
Hardware keys, software passwords, or any means needed to grant the hospital engineer full access to all service functionalities are the right of the hospital and should be delivered with the equipment, and should as well be valid throughout the whole lifetime of machine		

RM



طرح عطاء

يعلن مستشفى النجاح الوطني الجامعي عن طرح عطاء رقم RFQ-H-11735 والخاص بـ Fully Automated Blood Bank Analyzer على نظام ال Lease على الشركات المتخصصة والراغبة بالمشاركة بالعطاء تقديم عروض أسعارها في موعد أقصاه يوم الثلاثاء الموافق 22/10/2024 الساعة الثالثة مساءً PM 03:00 .

#	رقم العطاء	موضوع العطاء	نص نسخة العطاء	موعد الطرح	وقت الإغلاق	
					التاريخ	الساعة
1	RFQ-H-11735	Fully Automated Blood Bank Analyzer – Lease System	مجانياً	2024/10/14م	2024/10/22م	

حسب الشروط التالية:

1. ضرورة إرفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .

على الراغبين بالحصول على نسخة العطاء مراجعة دائرة اللوازم والمشتريات مكتب رقم 6400 / مبنى الإدارة أو عبر الإيبل Tender3@najah.edu وذلك خلال ساعات الروام الرسمية .
للاستفسار يرجى الاتصال على:

لأية استفسارات أخرى يرجى مراجعة دائرة اللوازم والمشتريات هاتف رقم 970-09-2331471 / تحويلة 6400 / فاكس: 970-9-2389685 +

مستشفى النجاح الوطني الجامعي