

التاريخ : 24/12/2025 م

طرح عطاء رقم – RFQ-H-11923

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين .

الموضوع : أجهزة طبية ، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكال الاسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .
2. ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء .
3. ضرورة أن تكون الأجهزة جديدة وليست مجددة .
4. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.
5. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، وفي حال تأخير التوريد مدة تزيد عن اسبوعين يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر مع تحميلكم فروق الأسعار .
6. الدفع خلال **90** يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .
7. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
8. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي بالظرف المختوم فقط وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح اي منهما العرض الفني والعرض المالي.
9. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .
10. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار .
11. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .
12. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
13. اخر موعد لتقديم عرض السعر هو يوم الأحد الموافق 2025-01-04 م الساعة الواحدة ظهراً 01:00 PM .
14. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (+97092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات

إياد المكاوي



12600868+12600869+12600878+12600875

التاريخ : 24/12/2025 م

طرح عطاء رقم – RFQ-H-11923

جدول الأسعار والكميات :

No .	NNUH Code	Item	QTY	Unit	CODE #	ويعمل / موزع	مدة التوريد	U.P (NIS)	T.P (NIS)
1	G-A-MDQ01249	Emergency Trolley	1	Pcs					
2	G-A-MDD00309	Defibrillator	1	Pcs					
3+4	G-A-MDD01276	Fiberoptic Laryngoscope Curved with All blades size 0,1,2,3,4	2	Pcs					
-	Total Price (NIS)								

يرجى تعبئة الجدول التالي مع ضرورة الختم على جميع أوراق عرض السعر .

1	إسم الشركة
2	الختم
3	التوقيع والإسم
4	معلومات الإتصال
5	مدة التوريد
6	مدة الكفالة إن وجدت

جدول المواصفات الفنية :



ITEM 3+4 Fiber optic laryngoscope, Qty: 1	Compliance Yes/No	Deviation
Good brand name with FDA or CE		
Manufacturer:		
Source and origin:		
Model:		
Fiber optic laryngoscope Specification:		
Durable metal		
Standard handle, diameter around 29mm		
Battery operated unit		
Lamp LED white colour		
Integrated fiber optic blades, Macintosh as attached picture		
With MAC blades size: 0,1,2,3 and 4		
With suitcase		
Autoclavable at 134°C		
Optional to be priced separately:		
1- List price all accessories necessary to operate the solution as per the required specs listed above		
2- List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
3- Lamp		
4- MAC blades size: 0,1,2,3 and 4		
Your offer should contain the following:		
1) CE or FDA and ISO certificates		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the product.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
Warranty: 1 year		



ITEM 2 Defibrillator	Compliance Yes/No	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Should have CB test and EMC test (60601.1 and 60601.2)		
Should has ISO 13485 certification for medical devices		
Manufacturer:		
Source and origin:		
Model:		
Defibrillator Specification:		
State of the art technology, the latest version, and the best technology from the manufacturer		
Biphasic, Manual and AED		
Heavy duty and durable machine		
Selectable energy with a wide range that suitable for adults, paediatric, and neonates resuscitation. (Please specify)		
Energy accuracy $\leq 10\%$		
Short charging time (please specify)		
Can perform defibrillation (manual), cardioversion, and pacing (for adults, pediatrics, neonates)		
Support AED mode with voice instructions and has pediatric/adult adhesive pads.		
7 inch color LCD approximately		
Automatic self test		
Should have adult and pediatric paddles integrated on same handle		
Momentary charge key on front panel and on the apex hand		
Monitor should display selected and delivered energy		
Sync message should display in case of selection of synchronization mode		
Long-life rechargeable battery (please specify the nominal number of uses for a fully charged battery?), more is better		
Multi-waveform and parameters visualization, (ECG 3-5-10 leads, SpO2, and RR in monitoring mode).		
Should have display for selected ECG input source (I, II, III, paddles)		
Lead off message should appear with alert tone		
Amplitude gain of ECG waveform should be adjustable		
User-friendly and easy to use interface		
Adjustable alarm limits and alarm mute/cancellation		
Should have alarm for high and low HR with a provision for alarm silence mode		
Ability to save events (please specify number of events capacity).		
Trend storage ≥ 48 hours.		
Should have an inbuilt thermal recorder		
Should supply along with pacing cable with pad		

Designed for easy and frequent cleaning/disinfection concept with hospital-grade products and compatible with disinfectant agents.		
AC input: 220V, 50Hz		
Should be supplied with all standard accessories		
Complete with all accessories and consumables needed to work completely as specified		
Training included:		
Operation training for the staff nursing team must be provided on-site		
Complete with the following accessories:		
1- ECG cable, 3-Lead cable clip tybe, Qty.2		
2- Battery, Qty.1		
3- Pacing cable with pad, Qty.1		
4- Spo2 sensor (neonatal , Qty.1), (pediatric, Qty.1) and (adult, Qty.1)		
Optional to be priced separately:		
List price all accessories necessary to operate the solution as per the required specs listed above		
List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
Screen		
All sensors		
Battery		
Power supply		
Your offer should contain the following:		
1) Preventive maintenance schedule		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the machine.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates and ISO 13485 certificates		
8) CB test and EMC test certificate		
9) The service and operational manuals must delivered as soft copy with the units		
10) Guarantee for 10 years for spare parts supply		
Warranty: 5 years, including the calibration, software upgrade, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.		
Preventive maintenance should be done as per the manufacturer's recommendations, all parts including PM kits to be replaced during the warranty period will be the responsibility of the supplier to provide on his own account. A schedule of the PM visits should be submitted to the hospital biomedical engineering dep.		
Hardware (service) keys, software passwords, website account or any means needed to grant the hospital engineer full access to all service functionalities are the right of the hospital and should be delivered with the equipment, and should as well be valid throughout the whole lifetime of machine.		

ITEM 1 Emergency Trolley	Compliance Yes/No	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Manufacturer:		
Source and origin:		
Model:		
Emergency Trolley Specification:		
State of the art technology and the latest version from the manufacturer and the latest and best technology on the market		
Material: Polymer		
Lock mechanism secures top compartment, drawers and side bins and can be sealed in segments or one seal secures all.		
Drawer dividers and trays		
Backboard can be mounted to the front or back of the cart.		
Include an extendable I.V. Pole, O2 Tank Storage, Suction Pump Shelf, Glove Box Holder, Lockable Sharps Container, Cord anagement, Trays & Dividers.		
One small drawer and 4 mediam drawers, 5 drawers in total		
Defibrillator swing arm		
Mobile on 4 castors with break mechanism		
Oxygen tank holder		
Drawer cassette for top cavity		
The trolley should be supplied with the following accessories:		
1- O2 cylinder 5 liter complete with regulator flowmeter and mask		
2- Diagnostic set		
3- Fully catheter 8 sizes 10 each		
4- Ambo bag set silicon neonatal, child, adult		
5- Dressing set (specify)		
6- Endotrachial tube 8 sizes 10 each		
7- Bag-Valve Masks (BVM): adult, pediatric and neonate sizes, with reservoir		
8- Oropharyngeal & nasopharyngeal airways, sizes 2–4 for adults and versions for pediatrics and neonatal		
9- Stylet, Magill forceps, bougie, ETT securing devices		

Complete with all accessories and consumables needed to work completely as specified		
Your offer should contain the following:		
1) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the product.		
2) Compliance sheet		
3) Catalogues		
4) Data sheet		
5) Authorization letter		
6) CE or FDA and ISO certificates		
7) The service and operational manuals must delivered as soft copy with the units		
8) Guarantee for 10 years for spare parts supply		
Warranty: 5 years		