التاريخ: 2026/11/03

طرح عطاء رقم RFQ-H-12028

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين

الموضوع: Portable ultrasound ، حسب المواصفات.

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على المواصفات المرفقة مع مراعاة الشروط التالية:

1. تكتب الأسعار **بالشيكل الإسرائيلي** وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل **قيمة الضريبة المضافة**، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد.
2. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل أو ما يعادلها، وفي حال عدم وجود خصم مصدر يتم خصم 10% من قيمة الإحالة.
3. أن يتضمن عرض السعر مدة صلاحية الاستخدام وعلى أن لا تقل عن سنتين.
4. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، ويحق للمستشفى إلغاء أمر الشراء والاحالة على مورد أخر دون تحمل اية مسؤولية تذكر في حالة تأخير التوريد ما يزيد عن أسبوعين.
5. تحديد فترة التوريد في عرض سعركم **وعدم ذلك يعني أن التوريد فوري** وايضاً أن تكون مدة صلاحية العرض المقدم من طرفكم لا تقل عن ثلاثة أشهر.
6. **تسلم عروض الأسعار بالظرف المغلق فقط**، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني.
7. **ضرورة الرد على العطاء سلباً أو إيجاباً، والالتزام بالتسعير على النموذج المرفق بالعطاء.**
8. يفضل أن تكون الأجهزة والمستهلكات الطبية صديقة للبيئة قدر الإمكان.
9. الأجهزة (الكهربائية/الطبية) إن وجدت:
   a. ألا يقل **تصنيف كفاءة** الطاقة عن **A** أو **+A** وفق نظام تصنيف كفاءة الطاقة المعمول به، وبما يعادل **كفاءة طاقة بنسبة 100%**
   b. أن تكون الأجهزة المعروضة AI Applicable مع تقديم ما يثبت ذلك رسمياً من الشركة المصنعة.
   c. ضرورة أن تكون الأجهزة المعروضة جديدة وليست مجددة.
10. الأدوية: أن يكون الدواء مسجلاً ومرخصاً وفقاً لشروط وزارة الصحة، وأن يتم التوريد والتسليم لصيدلية المستشفى وفقاً لتعليمات وزارة الصحة.
11. **الدفع حسب سياسات الدائرة المالية بالمستشفى.**
12. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار.
13. يحق للمستشفى إلغاء العطاء دون إبداء السبب ودون تحمل اية تكاليف مع إعادة رسوم نسخة العطاء للمشاركين في العطاء إن وجدت.
14. لا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل.
15. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر.
16. **آخر موعد لتقديم عرض السعر هو يوم الاربعاء الموافق 2026/11/11م الساعة العاشرة صباحاً AM (10:00).**
17. **لأية استفسارات متعلقة بالعطاء المذكور أعلاه، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (0097092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu فقط.**

مع فائق الاحترام،،،

أ.إياد مكاوي

مدير دائرة المشتريات والمستودعات

التاريخ: 2026/11/03

طرح عطاء رقم RFQ-H-12028

No	NNUH Code	Item	QTY	Unit	Brand/ Model	Unit Price (NIS)	Total Price (NIS)
1	G-A-MDD01829	Portable ultrasound	1	Pcs			
-	**Total Price (NIS)**						

Technical Specifications	Compliance (Yes / No)	Deviation / Bidder Comments
The offered medical device shall be supplied with valid **CE and/or FDA** certificate		
Should have CB test and EMC test (60601.1 and 60601.2)		
Should have ISO 13485 certification for medical devices		
Manufacturer:		
Source and origin:		
Model:		
Specifications:		
State of the art technology, the latest version, and the best technology from the manufacturer.		
Fully digital ultrasound system, compact portable ultrasound machine		
Heavy duty machine		
Color Doppler ultrasound for: Abdominal, obstetrical, small parts and general applications		
Automatic tissue optimization		
Imaging modes: B-Mode, M-Mode, Color M-Mode, Color Doppler, Power Doppler Imaging (PDI), PW Doppler, CW Doppler, Steerable Doppler, Tissue Harmonic Imaging (THI), Strain Imaging (SI), Strain Rate Imaging (SRI), 2D imaging, Color Flow Mode (CFM)		
High-definition (HD) digital black and white and colour liquid crystal display (LCD) monitor of at least 15 inches or more		
Display Resolution Not less than 1080 x 1280pixels		
Adjustable field of view		
Rechargeable Battery operated with battery life (run time) at least three hours		
Boot time less than 60 second		
The system must have archive capability for storage and retrieval of images and clips data		

Display Modes include: Duplex, Triplex, Split Screen, B/M & Dual B (B/B		
Harmonic Imaging available on all probes		
System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations.		
Cine memory on all modes		
Integrated speakers, audio volume adjustment		
PAN/ZOOM: Realtime image, frozen / stored image		
Post Processing technique on the Frozen and stored images		
Multi - format Report		
Frame Rate: Greater than 700 fps		
Image recording capability		
Contrast adjustment (gain control)		
Beam Steer: shifts beam into needle for better needle visualization		
Image grey scale: 256 shades of grey		
The system shall process a dynamic range that is at least 150db.		
With two ports for probe		
The Knobs and touch should be hybrid (can be used with gloves)and keyboard should fluid resistant		
Compatible with universal disinfectants		
The system must display at a maximum depth of 30cm		
The system and probes should be strong enough to survive from any type of collision or accidental drop. Please send any certifications proving drop test		
Data Transfer facility should be available as standard to transfer images		
Archiving:		
With integrated HDD (500GB), integrated DVD-R/W & 2 USB (2.0) connectors		
Integrated image and patient data management package		
Export in BMP, JPEG, TIFF, DICOM, MOV and AVI format		
Ability to export report in PDF format		
Printing reports and images directly to printer		
Ability to be connected to HIS and RIS		
The system shall support the all DICOM functionality, Storage, Print, and Work List, also ready to connect to PACS		
Power supply of 220-240 V, 50 Hz		
Complete with all accessories and consumables needed to work completely as specified		

TRAINING INCLUDED:		
On-site clinical operation training for laboratory/blood bank staff.		
On-site Biomedical Engineering training shall be provided for preventive maintenance, basic troubleshooting, configuration and system diagnostics.		

COMPLETE WITH THE FOLLOWING ACCESSORIES: (TO BE SUPPLIED WITH THE UNIT)		
Linear Probe 4 – 15 MHz		
Battery		

OPTIONAL TO BE PRICED SEPARATELY:		
List price all accessories necessary to operate the solution as per the required specs listed above		
List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
Convex Probe 1 – 5 MHz		
Linear Probe 4 – 15 MHz		
Phased array for adult		
Trolley		
Battery		

WARRANTY:		
Warranty: **Three years** (USA, EUR, or Jaban) brands, other equivalent brands with **five years** warranty Including the calibration, software upgrade, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.		
Preventive maintenance should be done as per the manufacturer's recommendations, all parts including PM kits to be replaced during the warranty period will be the responsibility of the supplier to provide on his own account. A schedule of the PM visits should be submitted to the hospital biomedical engineering department.		
Hardware (service) keys, software passwords, website account or any means needed to grant the hospital engineer full access to all service functionalities are the right of the hospital and should be delivered with the equipment, and should as well be valid throughout the whole lifetime of machine.		

- يرجى من حضرتكم تحديد مدة التوريد
- ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء، والالتزام بالعملة المحددة(شيكل).

1	اسم الشركة	
2	الختم	
3	الاسم والتوقيع	
4	معلومات الاتصال	
5	مدة التوريد	
6	مدة الكفالة	

إعلان طرح عطاء رقم RFQ-H-12028

- يعلن مستشفى النجاح الوطني الجامعي عن طرح عطاء (Portable ultrasound)
- على السادة الشركات المتخصصة والراغبة بالمشاركة بالعطاء تقديم عروض أسعارها في موعد أقصاه يوم الاربعاء الموافق 2026/11/11م الساعة العاشرة صباحاً am 10:00 .
- ضرورة إرفاق نسخة من الرخصة التجارية والسيرة الذاتية.
- في حال وجود أي استفسار يرجى التواصل مع دائرة اللوازم والمشتريات هاتف رقم 09-2389685 تحويلة رقم 6400 أو من خلال البريد الإلكتروني Tender3@najah.edu

#	رقم العطاء	موضوع العطاء	ثمن نسخة العطاء	كفالة دخول عطاء	موعد الطرح	موعد الإغلاق	
						التاريخ	الساعة
1	RFQ-H-12028	Portable ultrasound	مجاناً	-	2026/11/03	2026/11/11	am 10:00